

CERTIFICATE

Survey of 18 March 2022

You have fulfilled the requirements of the External Quality Assessment with the following analysis

Plasmaproteins 02 - Cystatin C (242):

Validity 12 months:

Cystatin C

(R) analysis is subject to the RiliBÄK

Participant:

59557

Ing. Eva Stupkova

Biovendor - Laboratorni medicina a.s.

Karasek 1767/1

62100 Brno

Düsseldorf, 31 March 2022



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. rer. nat. Ingo Schellenberg
(Adviser)

CERTIFICATE OF PARTICIPATION

Survey of 18 March 2022

You have participated in the External Quality Assessment with the following analysis

Plasmaproteins 02 - Cystatin C (242):

Cystatin C

(R) analysis is subject to the RiliBÄK

Participant:

59557


Ing. Eva Stupkova

Biovendor - Laboratorni medicina a.s.

Karasek 1767/1

62100 Brno

Düsseldorf, 31 March 2022



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. rer. nat. Ingo Schellenberg
(Adviser)

Listing and evaluation of the results

59557: Ing. Eva Stupkova
Biovendor - Laboratorni medicina a.s.



Survey of 18 March 2022

Adviser: Prof. Dr. rer. nat. Ingo Schellenberg
INSTAND e.V.
Ubierstraße 20
40223 Düsseldorf

Tel.: +49 211 1592 13 120
Mail: schellenberg@instand-ev.de

242

Plasmaproteins 02 - Cystatin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Cystatin C	21	mg/l	2.09	1.82	SV	1.37	2.28	14.6%	1,62	+	
	22		2.83	2.34	SV	1.76	2.93	20.9%	2,4	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Cystatin C	21	ng/ml	2085	0.001	64	ZY	ZY99
	22		2830	0.001			

Individual summary of results

59557: Ing. Eva Stupkova
Biovendor - Laboratorni medicina a.s.

Survey of 18 March 2022

242

Plasmaproteins 02 - Cystatin C

Cystatin C (mg/l, N = 46)										
Collective	Sample	Target value	Target range			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
Roche Diagnostics	21	1.45	1.09	-	1.81	1.45	7.52	16	93.8	93.8
	22	1.84	1.38	-	2.30	1.84	6.68	16	93.8	
other providers	21	1.82	1.37	-	2.28	1.82	9.01	30	96.7	96.7
	22	2.34	1.76	-	2.93	2.34	8.72	30	96.7	

Rate of success: 95,7%

