

CERTIFICATE OF PARTICIPATION

Survey of 15 March 2024

You have participated in the External Quality Assessment with the following analysis


Plasmaproteins 02 - Cystatin C (242):

Cystatin C

(R) analysis is subject to the RiliBÄK

Customer:
10059557
Ing. Eva Stupkova
Biovendor - Laboratorni medicina a.s.
Karasek 1767/1
62100 Brno

Düsseldorf, 24 April 2024



Prof. Dr. med. Michael Spannagl
(Head of Reference Institution)



Prof. Dr. rer. nat. Ingo Schellenberg
(Adviser)



Listing and evaluation of the results

10059557: Ing. Eva Stupkova
Biovendor - Laboratorni medicina a.s.

Survey of 15 March 2024

Adviser:

Prof. Dr. rer. nat. Ingo Schellenberg
IN STAND e.V.
Ubierstraße 20
40223 Düsseldorf

Tel.: +49 211 1592 13 120
Mail: schellenberg@instand-ev.de

242

Plasmaproteins 02 - Cystatin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Cystatin C	21	mg/l	1.47	1.11	SV	0.833	1.39	32.6%	2,97	-	
	22		2.11	1.59	SV	1.19	1.99	32.8%	3,58	-	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Cystatin C	21	ng/ml	1472	0.001	64	ZY	ZY99
	22		2112	0.001			

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".

Individual summary of results

10059557: Ing. Eva Stupkova
Biovendor - Laboratorni medicina a.s.

Survey of 15 March 2024

242

Plasmaproteins 02 - Cystatin C

Cystatin C (mg/l, N = 61)										
Collective	Sample	Target value	Target range		Participants collective			Rate (%)		
					AVG	CV	Num.	Sam.	total	
Siemens	21	1.42	1.07	-	1.78	1.42	12.1	22	100	95.5
	22	1.93	1.45	-	2.41	1.93	12.1	22	95.5	
Roche Diagnostics	21	0.974	0.730	-	1.22	0.974	4.00	24	100	100
	22	1.37	1.03	-	1.71	1.37	2.70	24	100	
other providers	21	1.11	0.833	-	1.39	1.11	11.0	15	93.3	93.3
	22	1.59	1.19	-	1.99	1.59	9.18	15	93.3	

Rate of success: 96,7%

