

ENG

Instructions for Use:
**TSH RECEPTOR AUTOANTIBODY
2.GENERATION ELISA**

Catalogue number:
RTRE/96/2AR

For research use only.

 **BioVendor**
R&D[®]



BioVendor – Laboratorní medicína a.s.

Karásek 1767/1, 621 00 Brno, Czech Republic

+420 549 124 185

info@biovendor.com

sales@biovendor.com

www.biovendor.com

1. INTENDED USE	3
2. ASSAY PRINCIPLE	3
3. STORAGE AND PREPARATION OF SERUM SAMPLES	3
4. MATERIALS REQUIRED AND NOT SUPPLIED	4
5. PREPARATION OF REAGENTS SUPPLIED	4
6. ASSAY PROCEDURE	6
7. RESULT ANALYSIS	7
8. TYPICAL RESULTS	7
9. ASSAY CUT OFF	8
10. CLINICAL EVALUATION	8
11. SAFETY CONSIDERATIONS	9
12. ASSAY PLAN	10
13. REFERENCES	11
14. EXPLANATION OF SYMBOLS	12

HISTORY OF CHANGES

Previous version	Current version
ENG.005A	ENG.006A
Symbol indicating the manufacturer added.	
History of changes added.	
Chapter 6, point 2.: running samples in duplicate recommended.	
Chapter 11.3. Diluent for SA-POD added.	

1. INTENDED USE

The TSH receptor (TSHR) autoantibody (TRAb) ELISA kit is intended for use by professional persons only for the quantitative determination of thyrotropin TRAb in human serum. Hyperthyroidism in Graves' disease is due to the presence of autoantibodies to the TSHR and measurement of these autoantibodies can be useful in disease diagnosis and management.

2. ASSAY PRINCIPLE

In TRAb ELISA, TRAb in patient sera, calibrators and controls are allowed to interact with TSHR coated onto ELISA plate wells. After a 2 hours incubation, the samples are discarded leaving TRAb bound to the immobilised TSHR. TSH biotin is added in a 2nd incubation step, where it interacts with immobilised TSHR, which have not been blocked by the bound TRAb from patient sera, calibrators or controls. The amount of TSH biotin bound to the plate is then determined in a 3rd incubation step by addition of streptavidin peroxidase (SA-POD), which binds specifically to biotin. Excess unbound SA-POD is then discarded and the addition of the peroxidase substrate 3,3',5,5' - tetramethylbenzidine (TMB) results in formation of a blue colour.

This reaction is stopped by the addition of stop solution causing the well contents to turn from blue to yellow. The absorbance of the yellow reaction mixture at 450 nm is then read using an ELISA plate reader. A lower absorbance indicates the presence of TRAb in the test sample as TRAb inhibits the binding of TSH biotin to TSHR coated plate wells. The measuring range is 1 – 40 U/l (NIBSC 08/204).

3. STORAGE AND PREPARATION OF SERUM SAMPLES

Sera to be analysed should be assayed soon after separation or stored, preferably in aliquots, at or below –20°C. 150 µl is sufficient for one assay (duplicate 75 µl determinations). Repeated freeze thawing or increases in storage temperature must be avoided. Incorrect storage of serum samples can lead to loss of TRAb activity. Do not use lipaemic or haemolysed serum samples. Do not use plasma in the assay. When required, thaw test sera at room temperature and mix gently to ensure homogeneity. Centrifuge the serum prior to assay (preferably for 5 minutes at 10-15000 rpm in a microfuge) to remove any particulate matter. **Please do not omit this centrifugation step for sera that are cloudy or contain particulates.**

4. MATERIALS REQUIRED AND NOT SUPPLIED

Pipettes capable of dispensing 50 µl, 75 µl and 100 µl.

Means of measuring out various volumes to reconstitute or dilute reagents.

Pure water.

ELISA Plate reader suitable for 96 well formats and capable of measuring at 450 nm.

ELISA Plate shaker, capable of 500 shakes/min (not an orbital shaker).

ELISA Plate cover.

5. PREPARATION OF REAGENTS SUPPLIED

Store unopened kits and all kit components (A-K) at 2–8 °C.

A TSH Receptor Coated Wells

12 breakapart strips of 8 wells (96 in total) in a frame and sealed in foil bag. Allow to stand at room temperature (20-25 °C) for at least 30 minutes before opening.

Ensure stripwells are fitted firmly into frame provided. After opening, return any unused wells to the original foil packet and seal, then place foil bag in the self-seal plastic bag with desiccant provided.

Store at 2-8 °C for up to expiry of kit.

B Start Buffer

10 ml

Ready for use

C1-4 Calibrators

1, 2, 8 and 40 U/l

(units are NIBSC 08/204)

4 x 1.0 ml

Ready for use

D1 Negative Control

1.0 ml

Ready for use

D2 Positive Control (See CoA for range)

1.0 ml

Ready for use

E TSH Biotin

3 x 4.5 ml

Lyophilised

Reconstitute each vial with 4.5 ml TSH biotin reconstitution buffer (F). When more than one vial is to be used, pool the vials and mix gently before use.

Store at 2–8 °C for up to expiry of kit.

F TSH Biotin Reconstitution Buffer

15 ml

Ready for use

G Streptavidin Peroxidase (SA-POD)

1 x 0.75 ml

Concentrated

Dilute 1 in 20 with diluent for SAPOD (H). For example, 0.5 ml (G) + 9.5 ml (H). Store

at 2–8°C for up to expiry of kit.

H Diluent for SA-POD

15 ml

Ready for use

I Peroxidase Substrate (TMB)

15 ml

Ready for use

J Concentrated Wash Solution

100 ml

Concentrated

Dilute to 1 litre with pure water before use. Store at 2–8°C up to expiry of kit.

K Stop Solution

10 ml

Ready for use

Example Version

6. ASSAY PROCEDURE

Allow all reagents and test samples to stand at room temperature (20-25°C) for at least 30 minutes before use. A repeating Eppendorf type pipette is recommended for steps 1, 5, 8, 10 and 11. Duplicate determinations are strongly recommended for test sera, calibrators and controls.

1. Pipette **75 µl** of start buffer (B) into each well to be used, leaving the last well for a blank (see step 12).
2. Pipette **75 µl** of patient sera, calibrators (C1-4) and controls (D1 and D2) into respective wells (start with the 40 U/l calibrator and descend down the plate to the negative control and then test sera, in duplicate recommended), leaving the last well blank.
3. Cover the frame and shake the wells for 2 hours at room temperature on an ELISA plate shaker (500 shakes per min.).
4. After incubation, aspirate samples by use of a plate washing machine, or discard the samples by briskly inverting the frame of wells over a suitable receptacle. Wash the wells once with diluted wash solution (J), and aspirate the wash by use of a plate washing machine or discard the wash by briskly inverting the frame of wells over a suitable receptacle. Tap the inverted wells gently on a clean, dry, absorbent surface to remove excess wash solution (only necessary if washing plate by hand).
5. Pipette **100 µl** of reconstituted TSH biotin (E) into each well (except blank). Avoid splashing the material out of the wells during addition.
6. Cover the frame, and incubate at room temperature for 25 minutes without shaking.
7. Repeat wash step 4.
8. Pipette **100 µl** of diluted SA-POD (G) into each well (except blank), cover the frame and incubate at room temperature for 20 minutes without shaking.
9. After incubation, aspirate samples by use of a plate washing machine, or discard the samples by briskly inverting the frame of stripwells over a suitable receptacle. Wash the wells twice with diluted wash solution (J) followed by once with pure water (to remove any foam) and tap the inverted wells gently on a clean, dry, absorbent surface to remove excess wash solution (if a plate washing machine is used, the plate can be washed 3 times with diluted wash solution (J) only).
10. Pipette **100 µl** of TMB (I) into each well (including blank) and incubate in the dark at room temperature for 30 minutes without shaking.
11. Pipette **50 µl** stop solution (K) to each well (including blank) and shake the plate for approximately 5 seconds on a plate shaker. Ensure substrate incubations are the same for each well.
12. Within 15 minutes, read the absorbance of each well at 450 nm using an ELISA plate reader, blanked against the well containing **100 µl** of TMB (I) and **50 µl** stop solution (K) only.

7. RESULT ANALYSIS

A calibration curve can be established by plotting calibrator concentration on the x-axis (log scale) against the absorbance of the calibrators on the y-axis (linear scale). The TRAb concentrations in patient sera can then be read off the calibration curve [plotted at BioVendor as a spline log/lin curve (smoothing factor=0)]. The negative control can be assigned a value of 0,1 to assist in computer processing of assay results. Other data reduction systems can be used. Results can also be expressed as inhibition (%) of TSH binding calculated using the formula


$$100 \times \left(1 - \frac{\text{test sample absorbance at 450 nm}}{\text{negative control (D1) absorbance 450 nm}} \right)$$

Samples with high TRAb concentrations can be diluted in kit negative control (D1). For example, 20 µl of sample plus 180 µl of negative control to give a 10x dilution. Other dilutions (e.g. 100x) can be prepared from a 10x dilution or otherwise as appropriate. Some sera will not dilute in a linear way and we suggest that the dilution giving a value closest to 50% inhibition is used for calculation of TRAb concentration.

8. TYPICAL RESULTS

Example only, not for calculation of actual results

Sample	A450 (minus blank)	%I	U/I
Control D1	2.00	0	0
C1	1.70	15	1
C2	1.50	25	2
C3	0.65	68	8
C4	0.15	93	40
Control D2	1.26	37	3.5



9. ASSAY CUT OFF

Cut off	U/I
Negative	≤ 1 U/I
Equivocal	1.1 – 1.5 U/I
Positive	> 1.5 U/I

This cut off has been validated at Biovendor. However each laboratory should establish its own normal and pathological reference ranges for TRAb levels. Also it is recommended that each laboratory include its own panel of control samples in the assay.

10. CLINICAL EVALUATION

10.1 Clinical Specificity

154 samples from healthy blood donors were assayed in the TRAb ELISA kit. 152 (99%) were identified as being negative for TRAb.

10.2 Clinical Sensitivity

50 samples from patients diagnosed with Graves' disease were assayed using the TRAb ELISA kit. 49 (98%) were identified as being positive for TSHb. 1 sample (2%) was identified as being within the equivocal range.

10.3 Functional Sensitivity

A plot of inter assay CV against U/I indicates a 20% CV occurring at 0.60 U/I.

10.4 Lower Detection Limit

The kit negative control was assayed 32 times and the mean and standard deviation calculated. The lower detection limit at 2 standard deviations was 0.21 U/I.

10.5 Inter Assay Precision

Sample	U/I (n=20)	CV (%)
1	3.9	12.9
2	5.4	10.9

10.6 Intra Assay Precision

Sample	U/I (n=25)	CV (%)
1	1.8	7.1
2	7.8	2.2

10.7 Clinical Accuracy

Analysis of sera from patients with autoimmune diseases other than Graves' disease indicated no interference from autoantibodies to thyroglobulin; thyroid peroxidase; glutamic acid decarboxylase; 21-hydroxylase; acetylcholine receptor; dsDNA or from rheumatoid factor.

10.8 Interference

No interference was observed when samples were spiked with the following materials; haemoglobin at 5 mg/ml; bilirubin at 0.2 mg/ml; Intralipid up to 30 mg/ml, human LH up to 10 u/ml; hCG up to 160 U/ml; human FSH up to 70 U/ml and human TSH up to 3 U/l.

The data quoted in these instructions should be used for guidance only. It is recommended that each laboratory include its own panel of control samples in the assay. Each laboratory should establish its own normal and pathological reference ranges for TRAb levels.

11. SAFETY CONSIDERATIONS

11.1 Streptavidin Peroxidase (SA-POD)



Signal word: Warning

Hazard statement(s)

H317: May cause an allergic skin reaction

Precautionary statement(s)

P280: Wear protective gloves/protective clothing/ eye protection/face protection

P302 + P352: IF ON SKIN: Wash with plenty of soap and water

P333 + P313: If skin irritation or rash occurs: Get medical advice/attention

P362 + P364: Take off contaminated clothing and wash it before reuse

11.2 Peroxidase Substrate (TMB)



Signal word: Danger

Hazard statement(s)

H360: May damage fertility or the unborn child

Precautionary statement(s)

P280: Wear protective gloves/protective clothing/ eye protection/face protection

P308 + P313: If exposed or concerned: Get medical advice/attention

11.3 Diluent for SA-POD and Concentrate Wash Solution

Hazard statement(s):

EUH208: Contains 2-Chloracetamide. May produce an allergic reaction.

This kit is intended for *in vitro* use by professional persons only. Follow the instructions carefully. Observe expiry dates stated on the labels and the specified stability for coated wells, diluted or reconstituted reagents. Refer to Materials Safety Data Sheet for more detailed safety information. Material of human origin used in the preparation of the kit has been tested and found non reactive for HIV1 and 2 and HCV antibodies and HBsAg but should, none the less, be handled as potentially infectious. Wash hands thoroughly if contamination has occurred and before leaving the laboratory. Sterilise all potentially contaminated waste, including test specimens before disposal. Material of animal origin used in the preparation of the kit has been obtained from animals certified as healthy but these materials should be handled as potentially infectious. Some components contain small quantities of sodium azide as preservative. With all kit components, avoid ingestion, inhalation, injection and contact with skin, eyes and clothing. Avoid formation of heavy metal azides in the drainage system by flushing any kit component away with copious amounts of water. including test specimens before disposal.

Material of animal origin used in the preparation of the kit has been obtained from animals certified as healthy but these materials should be handled as potentially infectious. Some components contain small quantities of sodium azide as preservative. With all kit components, avoid ingestion, inhalation, injection and contact with skin, eyes and clothing.

Avoid formation of heavy metal azides in the drainage system by flushing any kit component away with copious amounts of water.

12. ASSAY PLAN










Allow all reagents and samples to reach room temperature (20-25 °C) before use.	
Pipette:	75 µl Start buffer into each well (except blank)
Pipette:	75 µl Calibrators (starting with the highest concentration and descending to lowest), kit controls, patient sera (except blank)
Incubate	2 hours at room temperature on an ELISA plate shaker at 500 shakes/min
Aspirate/Decant:	Plate
Wash:	Plate once on automatic washer (or wash once, invert and tap dry on absorbent material for manual washing)
Pipette:	100 µl TSH biotin (reconstituted) into each well (except blank)
Incubate	25 minutes at room temperature without shaking
Aspirate/Decant:	Plate
Wash:	Plate once as above
Pipette:	100 µl SAPOD (diluted 1:20) into each well (except blank)
Incubate	20 minutes at room temperature without shaking
Aspirate/Decant:	Plate
Wash:	Plate three times on automatic washer (or wash twice, rinse once with pure water and dry on absorbent material for manual washing)
Pipette:	100 µl TMB into each well (including blank)
Incubate:	30 minutes at room temperature in the dark without shaking
Pipette:	50 µl stop solution into each well (including blank) and shake for 5 seconds
Read absorbance at 450 nm, within 15 minutes of adding stop solution	
Do not perform the assay at temperatures above 25 °C.	

13. REFERENCES

1. J. Bolton et al Measurement of thyroid stimulating hormone receptor autoantibodies by ELISA Clin. Chem 1999 45: 2285-2287
2. K. Kamijo TSH receptor antibody measurement in patients with various thyrotoxicosis and Hashimoto's thyroiditis: a comparison of two two-step assays, coated plate ELISA using porcine TSH receptor and coated tube radioassay using human recombinant TSH receptor Endocrine Journal 2003 50:113-116
3. B. Rees Smith et al A new assay for thyrotropin receptor autoantibodies Thyroid 2004 14: 830-835

Example Version

14. EXPLANATION OF SYMBOLS

	Catalogue number
	Batch code
	Caution
	Use by date
	Temperature limit
	Manufacturer
 www.biovendor.com	Read electronic instructions for use - eIFU
	The content is sufficient for 96 tests
	Biological risks



BioVendor – Laboratorní medicína a.s.

Karásek 1767/1, 621 00 Brno, Czech Republic

+420 549 124 185

info@biovendor.com

sales@biovendor.com

www.biovendor.com

